

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW122821\
 Data File : VW024218.D
 Acq On : 28 Dec 2021 15:12
 Operator : SY/MD
 Sample : M5201-12
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 P077-TW001-01

Quant Time: Dec 29 05:05:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR122721WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Dec 29 04:43:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	72561	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	75149	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	37821	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	25895	4.745	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.800%
7) Chloroethane-d5	1.568	69	23801	5.229	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.600%
11) 1,1-Dichloroethene-d2	2.108	63	40965	3.088	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	61.800%
20) 2-Butanone-d5	3.899	46	53641	88.480	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	176.960%#
24) Chloroform-d	4.349	84	60567	5.228	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.600%
26) 1,2-Dichloroethane-d4	5.034	65	30849	5.948	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	119.000%
32) Benzene-d6	5.053	84	110263	5.433	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.600%
36) 1,2-Dichloropropane-d6	6.069	67	31006	5.578	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	111.600%
41) Toluene-d8	7.317	98	97850	4.892	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.800%
43) trans-1,3-Dichloroprop...	7.625	79	11586	4.922	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	98.400%
46) 2-Hexanone-d5	8.088	63	42134	53.236	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	106.480%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	21952	5.784	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	115.600%
66) 1,2-Dichlorobenzene-d4	11.625	152	39152	5.955	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	119.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV122821\
Data File : VV024218.D
Acq On : 28 Dec 2021 15:12
Operator : SY/MD
Sample : M5201-12
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
P077-TW001-01

Quant Time: Dec 29 05:05:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR122721WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Dec 29 04:43:10 2021
Response via : Initial Calibration

