

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV122920\
 Data File : VV019865.D
 Acq On : 29 Dec 2020 15:40
 Operator : SY/MD
 Sample : VV1229WBL01
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK301

Quant Time: Dec 30 05:04:50 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVTR122920WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Dec 30 04:48:15 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.62	114	285263	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.86	117	263487	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.26	152	125235	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.30	65	99187	4.83	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	96.60%
7) Chloroethane-d5	1.56	69	83684	4.89	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.80%
11) 1,1-Dichloroethene-d2	2.10	63	154348	3.63	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.60%
20) 2-Butanone-d5	3.93	46	207379	47.88	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.76%
24) Chloroform-d	4.35	84	172379	4.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.80%
26) 1,2-Dichloroethane-d4	5.04	65	91676	4.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.80%
32) Benzene-d6	5.05	84	358862	4.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.20%
36) 1,2-Dichloropropane-d6	6.07	67	107901	4.97	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.40%
41) Toluene-d8	7.32	98	306319	4.86	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.20%
43) trans-1,3-Dichloropropene-	7.63	79	34752	4.59	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.80%
46) 2-Hexanone-d5	8.10	63	132513	43.53	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	87.06%
56) 1,1,2,2-Tetrachloroethane-	10.22	84	59524	4.38	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.60%
66) 1,2-Dichlorobenzene-d4	11.63	152	96273	4.95	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.00%

Target Compounds					Ovalue
16) Methylene chloride	2.50	84	2651	0.110 ug/L	95
25) Chloroform	4.39	83	8565	0.199 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV122920\
Data File : VV019865.D
Acq On : 29 Dec 2020 15:40
Operator : SY/MD
Sample : VV1229WBL01
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK301

Quant Time: Dec 30 05:04:50 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVTR122920WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Dec 30 04:48:15 2020
Response via : Initial Calibration

