

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV123019\
 Data File : VV014224.D
 Acq On : 30 Dec 2019 17:32
 Operator : SY/MD
 Sample : K6459-08
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H0BP7

Quant Time: Dec 31 06:42:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR122719WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Dec 31 06:00:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	871516	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	789575	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	368575	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	217216	5.32	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	106.40%
7) Chloroethane-d5	1.59	69	193014	5.68	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	113.60%
11) 1,1-Dichloroethene-d2	2.13	63	277094	3.78	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.60%
20) 2-Butanone-d5	3.93	46	608482	60.01	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	120.02%
24) Chloroform-d	4.40	84	432048	4.22	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.40%
26) 1,2-Dichloroethane-d4	5.08	65	245243	4.97	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.40%
32) Benzene-d6	5.10	84	1004400	4.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.20%
36) 1,2-Dichloropropane-d6	6.12	67	314750	4.82	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.40%
41) Toluene-d8	7.36	98	912561	4.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.60%
43) trans-1,3-Dichloropropene-	7.66	79	131190	4.49	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.80%
46) 2-Hexanone-d5	8.13	63	532557	48.07	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.14%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	225240	4.83	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.60%
64) 1,2-Dichlorobenzene-d4	11.67	152	305098	4.62	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.40%

Target Compounds					Ovalue
25) Chloroform	4.42	83	58352	0.516 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV123019\
Data File : VV014224.D
Acq On : 30 Dec 2019 17:32
Operator : SY/MD
Sample : K6459-08
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H0BP7

Quant Time: Dec 31 06:42:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR122719WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Dec 31 06:00:54 2019
Response via : Initial Calibration

