

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV123020\
 Data File : VV019872.D
 Acq On : 30 Dec 2020 11:00
 Operator : SY/MD
 Sample : VV1230WBL04
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK302

Quant Time: Dec 31 02:09:12 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVTR122920WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Dec 31 02:03:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.62	114	271387	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.86	117	257185	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.26	152	117676	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.30	65	91139	4.67	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.40%
7) Chloroethane-d5	1.56	69	77866	4.78	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.60%
11) 1,1-Dichloroethene-d2	2.10	63	144705	3.58	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.60%
20) 2-Butanone-d5	3.93	46	194562	47.22	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	94.44%
24) Chloroform-d	4.35	84	168563	4.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.20%
26) 1,2-Dichloroethane-d4	5.04	65	87995	4.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.60%
32) Benzene-d6	5.05	84	338987	4.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.00%
36) 1,2-Dichloropropane-d6	6.07	67	102150	4.82	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.40%
41) Toluene-d8	7.32	98	288401	4.68	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.60%
43) trans-1,3-Dichloropropene-	7.63	79	32887	4.45	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.00%
46) 2-Hexanone-d5	8.10	63	124615	41.94	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.88%
56) 1,1,2,2-Tetrachloroethane-	10.22	84	57851	4.37	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.40%
66) 1,2-Dichlorobenzene-d4	11.63	152	91594	5.01	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV123020\
Data File : VV019872.D
Acq On : 30 Dec 2020 11:00
Operator : SY/MD
Sample : VV1230WBL04
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK302

Quant Time: Dec 31 02:09:12 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVTR122920WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Dec 31 02:03:35 2020
Response via : Initial Calibration

