

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV123119\
 Data File : VV014231.D
 Acq On : 31 Dec 2019 11:44
 Operator : SY/MD
 Sample : VV1231WBL01
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK43

Quant Time: Jan 01 10:33:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR122719WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jan 01 03:03:23 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	871268	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	803191	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	367089	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	172181	4.22	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.40%
7) Chloroethane-d5	1.58	69	155239	4.57	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.40%
11) 1,1-Dichloroethene-d2	2.13	63	241723	3.30	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.00%
20) 2-Butanone-d5	3.95	46	522877	51.58	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.16%
24) Chloroform-d	4.40	84	442432	4.32	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.40%
26) 1,2-Dichloroethane-d4	5.08	65	240191	4.87	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.40%
32) Benzene-d6	5.09	84	977556	4.45	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.00%
36) 1,2-Dichloropropane-d6	6.11	67	313403	4.72	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.40%
41) Toluene-d8	7.35	98	892157	4.35	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.00%
43) trans-1,3-Dichloropropene-	7.66	79	125081	4.21	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.20%
46) 2-Hexanone-d5	8.13	63	549470	48.75	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.50%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	217547	4.59	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.80%
64) 1,2-Dichlorobenzene-d4	11.67	152	298473	4.54	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	90.80%

Target Compounds					Ovalue
16) Methylene chloride	2.53	84	8933	0.138 ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV123119\
Data File : VV014231.D
Acq On : 31 Dec 2019 11:44
Operator : SY/MD
Sample : VV1231WBL01
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK43

Quant Time: Jan 01 10:33:23 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR122719WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jan 01 03:03:23 2020
Response via : Initial Calibration

