

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV123119\
 Data File : VV014238.D
 Acq On : 31 Dec 2019 15:04
 Operator : SY/MD
 Sample : K6459-23
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H0BR1

Quant Time: Jan 02 11:28:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR122719WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jan 01 03:03:23 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	749180	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	696797	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	321509	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	176442	5.02	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.40%
7) Chloroethane-d5	1.58	69	174267	5.96	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	119.20%
11) 1,1-Dichloroethene-d2	2.13	63	257743	4.09	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	81.80%
20) 2-Butanone-d5	3.92	46	636554	73.03	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	146.06%#
24) Chloroform-d	4.40	84	416408	4.73	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.60%
26) 1,2-Dichloroethane-d4	5.08	65	240787	5.67	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	113.40%
32) Benzene-d6	5.10	84	984513	5.17	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.40%
36) 1,2-Dichloropropane-d6	6.11	67	317201	5.51	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	110.20%
41) Toluene-d8	7.36	98	892081	5.02	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.40%
43) trans-1,3-Dichloropropene-	7.66	79	126740	4.92	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	98.40%
46) 2-Hexanone-d5	8.13	63	559238	57.20	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	114.40%
57) 1,1,2,2-Tetrachloroethane-	10.25	84	223065	5.42	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	108.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	303426	5.27	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV123119\
Data File : VV014238.D
Acq On : 31 Dec 2019 15:04
Operator : SY/MD
Sample : K6459-23
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H0BR1

Quant Time: Jan 02 11:28:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR122719WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jan 01 03:03:23 2020
Response via : Initial Calibration

