

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV123120\
 Data File : VV019879.D
 Acq On : 31 Dec 2020 14:05
 Operator : SY/MD
 Sample : VV1231WBL06
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK306

Quant Time: Jan 01 00:47:21 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVTR122920WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jan 01 00:38:14 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.62	114	259284	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.86	117	243975	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.26	152	113949	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.30	65	82055	4.40	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.00%
7) Chloroethane-d5	1.56	69	73009	4.69	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.80%
11) 1,1-Dichloroethene-d2	2.10	63	131985	3.42	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	68.40%
20) 2-Butanone-d5	3.94	46	183184	46.54	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	93.08%
24) Chloroform-d	4.35	84	150521	4.36	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.20%
26) 1,2-Dichloroethane-d4	5.04	65	80161	4.61	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.20%
32) Benzene-d6	5.05	84	309986	4.63	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.60%
36) 1,2-Dichloropropane-d6	6.07	67	95327	4.74	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.80%
41) Toluene-d8	7.32	98	261177	4.47	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.40%
43) trans-1,3-Dichloropropene-	7.63	79	28806	4.11	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.20%
46) 2-Hexanone-d5	8.10	63	117694	41.76	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.52%
56) 1,1,2,2-Tetrachloroethane-	10.22	84	51822	4.12	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	82.40%
66) 1,2-Dichlorobenzene-d4	11.63	152	83368	4.71	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.20%

Target Compounds					Ovalue
16) Methylene chloride	2.50	84	7182	0.328 ug/L	95
25) Chloroform	4.38	83	10716	0.274 ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV123120\
Data File : VV019879.D
Acq On : 31 Dec 2020 14:05
Operator : SY/MD
Sample : VV1231WBL06
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK306

Quant Time: Jan 01 00:47:21 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVTR122920WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jan 01 00:38:14 2021
Response via : Initial Calibration

