

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV123120\
 Data File : VV019880.D
 Acq On : 31 Dec 2020 14:28
 Operator : SY/MD
 Sample : VV1231WBL07
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK307

Quant Time: Jan 01 00:53:40 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVTR122920WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jan 01 00:38:14 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.62	114	283781	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.86	117	273015	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.26	152	124559	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.30	65	88442	4.33	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	86.60%
7) Chloroethane-d5	1.57	69	77624	4.55	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.00%
11) 1,1-Dichloroethene-d2	2.10	63	138469	3.27	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.40%
20) 2-Butanone-d5	3.94	46	186307	43.24	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	86.48%
24) Chloroform-d	4.35	84	159775	4.23	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.60%
26) 1,2-Dichloroethane-d4	5.04	65	87306	4.58	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.60%
32) Benzene-d6	5.05	84	330953	4.42	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.40%
36) 1,2-Dichloropropane-d6	6.08	67	102446	4.55	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.00%
41) Toluene-d8	7.32	98	279787	4.28	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.60%
43) trans-1,3-Dichloropropene-	7.63	79	33010	4.21	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.20%
46) 2-Hexanone-d5	8.10	63	127551	40.44	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	80.88%
56) 1,1,2,2-Tetrachloroethane-	10.23	84	56693	4.03	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	80.60%
66) 1,2-Dichlorobenzene-d4	11.63	152	89601	4.63	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.60%

Target Compounds					Ovalue
16) Methylene chloride	2.51	84	3305	0.138 ug/L	91
25) Chloroform	4.38	83	13801	0.323 ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV123120\
Data File : VV019880.D
Acq On : 31 Dec 2020 14:28
Operator : SY/MD
Sample : VV1231WBL07
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK307

Quant Time: Jan 01 00:53:40 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SFAMVTR122920WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jan 01 00:38:14 2021
Response via : Initial Calibration

