

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW123121\
 Data File : VW024251.D
 Acq On : 30 Dec 2021 23:33
 Operator : SY/MD
 Sample : M5194-22
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK002

Quant Time: Dec 31 06:17:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM123121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 31 06:12:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	247932	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	231521	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	115021	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	85738	48.050	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	96.100%
7) Chloroethane-d5	1.568	69	64310	47.444	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	94.880%
11) 1,1-Dichloroethene-d2	2.108	63	123279	36.967	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.940%
21) 2-Butanone-d5	3.886	46	89053	89.579	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	89.580%
24) Chloroform-d	4.346	84	148568	45.557	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.120%
26) 1,2-Dichloroethane-d4	5.027	65	101975	46.704	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.400%
32) Benzene-d6	5.047	84	291733	48.287	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.580%
36) 1,2-Dichloropropane-d6	6.066	67	79176	46.761	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.520%
41) Toluene-d8	7.313	98	272553	46.593	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.180%
43) trans-1,3-Dichloroprop...	7.619	79	45259	46.419	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.840%
47) 2-Hexanone-d5	8.085	63	66414	90.282	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	90.280%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	104704	43.271	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.540%
66) 1,2-Dichlorobenzene-d4	11.625	152	114121	49.777	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.560%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV123121\
Data File : VV024251.D
Acq On : 30 Dec 2021 23:33
Operator : SY/MD
Sample : M5194-22
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK002

Quant Time: Dec 31 06:17:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM123121WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 31 06:12:03 2021
Response via : Initial Calibration

