

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW011624\
 Data File : VW033660.D
 Acq On : 16 Jan 2024 11:01
 Operator : SY/MD
 Sample : P1131-03
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AY5

Quant Time: Jan 17 00:10:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011224WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jan 17 00:08:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	115872	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	107576	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	54070	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	24387	3.849	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	77.000%
7) Chloroethane-d5	1.536	69	25141	4.474	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.400%
11) 1,1-Dichloroethene-d2	2.063	65	12335	4.320	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	86.400%
20) 2-Butanone-d5	3.806	46	71229	45.978	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	91.960%
24) Chloroform-d	4.253	84	63111	4.606	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.200%
26) 1,2-Dichloroethane-d4	4.947	65	30804	5.014	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.200%
32) Benzene-d6	4.963	84	120544	4.644	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.800%
36) 1,2-Dichloropropane-d6	5.992	67	39765	4.595	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.000%
41) Toluene-d8	7.246	98	108168	4.668	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.400%
43) trans-1,3-Dichloroprop...	7.561	79	13467	4.501	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.000%
46) 2-Hexanone-d5	8.027	63	55663	47.388	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.780%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	26163	5.174	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.400%
66) 1,2-Dichlorobenzene-d4	11.561	152	40925	4.930	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011624\
Data File : VV033660.D
Acq On : 16 Jan 2024 11:01
Operator : SY/MD
Sample : P1131-03
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AY5

Quant Time: Jan 17 00:10:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011224WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jan 17 00:08:33 2024
Response via : Initial Calibration

