

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW011624\
 Data File : VW033662.D
 Acq On : 16 Jan 2024 11:49
 Operator : SY/MD
 Sample : P1131-05
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AY7

Quant Time: Jan 17 00:10:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011224WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jan 17 00:08:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.538	114	114903	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	104391	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	53583	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	22656	3.606	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	72.200%	
7) Chloroethane-d5	1.535	69	23587	4.233	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	84.600%	
11) 1,1-Dichloroethene-d2	2.063	65	11239	3.969	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	79.400%	
20) 2-Butanone-d5	3.812	46	71799	46.737	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	93.480%	
24) Chloroform-d	4.252	84	59870	4.406	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	88.200%	
26) 1,2-Dichloroethane-d4	4.950	65	30077	4.937	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	98.800%	
32) Benzene-d6	4.963	84	114172	4.533	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	90.600%	
36) 1,2-Dichloropropane-d6	5.992	67	37508	4.466	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	89.400%	
41) Toluene-d8	7.246	98	98532	4.382	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	87.600%	
43) trans-1,3-Dichloroprop...	7.561	79	12484	4.300	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	86.000%	
46) 2-Hexanone-d5	8.030	63	51948	45.574	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	91.140%	
56) 1,1,2,2-Tetrachloroeth...	10.156	84	23259	4.740	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	94.800%	
66) 1,2-Dichlorobenzene-d4	11.561	152	38314	4.657	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	93.200%	
Target Compounds						
					Qvalue	
34) Trichloroethene	5.847	95	2844	0.288	ug/L	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011624\
Data File : VV033662.D
Acq On : 16 Jan 2024 11:49
Operator : SY/MD
Sample : P1131-05
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AY7

Quant Time: Jan 17 00:10:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011224WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jan 17 00:08:33 2024
Response via : Initial Calibration

