

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW011624\
 Data File : VW033699.D
 Acq On : 17 Jan 2024 02:55
 Operator : SY/MD
 Sample : P1143-05
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AZ0

Quant Time: Jan 17 03:34:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011224WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jan 17 01:35:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	107593	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	100418	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	52603	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	19087	3.244	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	64.800%	
7) Chloroethane-d5	1.536	69	19667	3.769	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	75.400%	
11) 1,1-Dichloroethene-d2	2.063	65	9857	3.718	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	74.400%	
20) 2-Butanone-d5	3.812	46	64735	45.002	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	90.000%	
24) Chloroform-d	4.249	84	53593	4.212	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	84.200%	
26) 1,2-Dichloroethane-d4	4.947	65	26147	4.583	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	91.600%	
32) Benzene-d6	4.963	84	97621	4.029	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	80.600%	
36) 1,2-Dichloropropane-d6	5.992	67	33163	4.105	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	82.200%	
41) Toluene-d8	7.246	98	88785	4.105	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	82.000%	
43) trans-1,3-Dichloroprop...	7.561	79	11307	4.049	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	81.000%	
46) 2-Hexanone-d5	8.030	63	53625	48.907	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	97.820%	
56) 1,1,2,2-Tetrachloroeth...	10.153	84	22570	4.782	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	95.600%	
66) 1,2-Dichlorobenzene-d4	11.561	152	34822	4.312	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	86.200%	
Target Compounds						
22) cis-1,2-Dichloroethene	3.822	96	2605	0.290	ug/L	83
29) 1,1,1-Trichloroethane	4.516	97	1413	0.099	ug/L #	84
34) Trichloroethene	5.847	95	2787	0.294	ug/L	84
47) Tetrachloroethene	7.908	164	2899	0.425	ug/L	93
63) 1,2,4-Trimethylbenzene	10.857	105	2875	0.079	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011624\
Data File : VV033699.D
Acq On : 17 Jan 2024 02:55
Operator : SY/MD
Sample : P1143-05
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AZ0

Quant Time: Jan 17 03:34:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011224WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jan 17 01:35:32 2024
Response via : Initial Calibration

