

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW011624\
 Data File : VW033658.D
 Acq On : 16 Jan 2024 10:14
 Operator : SY/MD
 Sample : P1131-01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AY2

Quant Time: Jan 17 00:09:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011224WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jan 17 00:08:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	107849	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	99652	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	48887	5.000	ug/L	0.00

System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	23831	4.041	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.800%
7) Chloroethane-d5	1.536	69	24418	4.669	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.400%
11) 1,1-Dichloroethene-d2	2.060	65	11824	4.449	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	89.000%
20) 2-Butanone-d5	3.825	46	66081	45.828	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	91.660%
24) Chloroform-d	4.249	84	62503	4.901	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.000%
26) 1,2-Dichloroethane-d4	4.944	65	29476	5.155	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.000%
32) Benzene-d6	4.960	84	116853	4.860	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.200%
36) 1,2-Dichloropropane-d6	5.989	67	37273	4.650	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.000%
41) Toluene-d8	7.246	98	105483	4.914	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.200%
43) trans-1,3-Dichloroprop...	7.561	79	12808	4.621	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.400%
46) 2-Hexanone-d5	8.031	63	51987	47.778	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.560%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	24145	5.155	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.000%
66) 1,2-Dichlorobenzene-d4	11.561	152	38111	5.077	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011624\
Data File : VV033658.D
Acq On : 16 Jan 2024 10:14
Operator : SY/MD
Sample : P1131-01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AY2

Quant Time: Jan 17 00:09:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011224WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jan 17 00:08:33 2024
Response via : Initial Calibration

