

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW011624\
 Data File : VW033666.D
 Acq On : 16 Jan 2024 13:24
 Operator : SY/MD
 Sample : P1062-12
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BH265

Quant Time: Jan 17 00:11:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011224WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jan 17 00:08:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	111469	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	102100	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	52259	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	22089	3.624	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	72.400%
7) Chloroethane-d5	1.536	69	22944	4.244	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	84.800%
11) 1,1-Dichloroethene-d2	2.060	65	11646	4.240	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	84.800%
20) 2-Butanone-d5	3.819	46	66867	44.867	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	89.740%
24) Chloroform-d	4.253	84	59469	4.512	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.200%
26) 1,2-Dichloroethane-d4	4.947	65	28418	4.808	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.200%
32) Benzene-d6	4.963	84	112144	4.552	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.000%
36) 1,2-Dichloropropane-d6	5.992	67	36098	4.395	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.000%
41) Toluene-d8	7.246	98	98689	4.487	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.800%
43) trans-1,3-Dichloroprop...	7.561	79	11305	3.981	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.600%
46) 2-Hexanone-d5	8.031	63	52600	47.182	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.360%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	23973	4.995	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.000%
66) 1,2-Dichlorobenzene-d4	11.561	152	37545	4.679	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV011624\
Data File : VV033666.D
Acq On : 16 Jan 2024 13:24
Operator : SY/MD
Sample : P1062-12
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH265

Quant Time: Jan 17 00:11:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011224WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jan 17 00:08:33 2024
Response via : Initial Calibration

