

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050922\
 Data File : VV025852.D
 Acq On : 09 May 2022 10:47
 Operator : SY/MD
 Sample : VV0509WBL01
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK358

Quant Time: May 10 04:47:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR050322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue May 10 04:47:02 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	193874	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	195734	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	85196	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	74052	3.485	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	69.600%
7) Chloroethane-d5	1.568	69	78685	4.210	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	84.200%
11) 1,1-Dichloroethene-d2	2.105	63	127678	3.163	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	63.200%
20) 2-Butanone-d5	3.899	46	145870	50.158	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.320%
24) Chloroform-d	4.342	84	147618	4.638	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.800%
26) 1,2-Dichloroethane-d4	5.027	65	68990	4.897	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.000%
32) Benzene-d6	5.047	84	282497	4.494	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.800%
36) 1,2-Dichloropropane-d6	6.066	67	87900	4.755	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.000%
41) Toluene-d8	7.313	98	228024	4.010	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.200%
43) trans-1,3-Dichloroprop...	7.619	79	24500	4.186	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.800%
46) 2-Hexanone-d5	8.088	63	118449	51.934	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.860%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	57120	4.894	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.800%
66) 1,2-Dichlorobenzene-d4	11.622	152	77375	5.178	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV050922\
Data File : VV025852.D
Acq On : 09 May 2022 10:47
Operator : SY/MD
Sample : VV0509WBL01
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK358

Quant Time: May 10 04:47:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR050322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue May 10 04:47:02 2022
Response via : Initial Calibration

