

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW052022\
 Data File : VW025942.D
 Acq On : 20 May 2022 13:44
 Operator : SY/MD
 Sample : N2955-02
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AA1

Quant Time: May 20 21:50:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR051322WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri May 20 21:49:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	157930	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	164378	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	74336	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	59179	4.537	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.800%
7) Chloroethane-d5	1.571	69	58780	5.263	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	105.200%
11) 1,1-Dichloroethene-d2	2.111	63	97993	3.394	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.800%
20) 2-Butanone-d5	3.921	46	82643	54.711	ug/L	0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.420%
24) Chloroform-d	4.352	84	122176	5.735	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	114.600%
26) 1,2-Dichloroethane-d4	5.037	65	52358	5.795	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	115.800%
32) Benzene-d6	5.053	84	229334	5.373	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.400%
36) 1,2-Dichloropropane-d6	6.072	67	72179	5.816	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	116.400%
41) Toluene-d8	7.317	98	184721	4.655	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.200%
43) trans-1,3-Dichloroprop...	7.628	79	16828	4.146	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.000%
46) 2-Hexanone-d5	8.091	63	77433	51.387	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.780%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	45491	5.691	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	113.800%
66) 1,2-Dichlorobenzene-d4	11.622	152	65998	6.236	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	124.800%#
Target Compounds						Qvalue
3) Chloromethane	1.243	50	3588	0.180	ug/L	86
25) Chloroform	4.384	83	8324	0.328	ug/L	96
47) Tetrachloroethene	7.979	164	1340	0.122	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052022\
Data File : VV025942.D
Acq On : 20 May 2022 13:44
Operator : SY/MD
Sample : N2955-02
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AA1

Quant Time: May 20 21:50:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR051322WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri May 20 21:49:46 2022
Response via : Initial Calibration

