

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW052622\
 Data File : VW025956.D
 Acq On : 26 May 2022 14:20
 Operator : SY/MD
 Sample : VW0526WBL01
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK215

Quant Time: May 27 01:28:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri May 27 01:28:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	351541	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	343294	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	158868	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.310	65	194421	42.853	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	85.700%
7) Chloroethane-d5	1.571	69	155429	39.011	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	78.020%
11) 1,1-Dichloroethene-d2	2.108	63	272412	34.330	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.660%
21) 2-Butanone-d5	3.899	46	152219	78.278	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	78.280%
24) Chloroform-d	4.346	84	275720	43.510	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.020%
26) 1,2-Dichloroethane-d4	5.027	65	171056	44.664	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.320%
32) Benzene-d6	5.047	84	565309	47.236	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.480%
36) 1,2-Dichloropropane-d6	6.066	67	177651	48.330	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.660%
41) Toluene-d8	7.313	98	485723	44.050	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.100%
43) trans-1,3-Dichloroprop...	7.619	79	66722	40.664	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.320%
47) 2-Hexanone-d5	8.088	63	101732	75.490	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	75.490%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	216599	44.187	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	88.380%
66) 1,2-Dichlorobenzene-d4	11.622	152	174592	50.100	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052622\
Data File : VV025956.D
Acq On : 26 May 2022 14:20
Operator : SY/MD
Sample : VV0526WBL01
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK215

Quant Time: May 27 01:28:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051022WMA.M
Quant Title : VOC Analysis
QLast Update : Fri May 27 01:28:09 2022
Response via : Initial Calibration

