

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW052622\
 Data File : VW025960.D
 Acq On : 26 May 2022 15:59
 Operator : SY/MD
 Sample : VIBLK271
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK271

Quant Time: May 27 01:29:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri May 27 01:28:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	76956	50.000	ug/L	0.01
28) Chlorobenzene-d5	8.854	117	80009	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	33814	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	61105	61.524	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	= 123.040%	
7) Chloroethane-d5	1.571	69	48361	55.447	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	= 110.900%	
11) 1,1-Dichloroethene-d2	2.111	63	80514	46.351	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 92.700%	
21) 2-Butanone-d5	3.915	46	33907	79.651	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	= 79.650%	
24) Chloroform-d	4.352	84	82151	59.220	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 118.440%	
26) 1,2-Dichloroethane-d4	5.037	65	52584	62.720	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 125.440%#	
32) Benzene-d6	5.053	84	165558	59.356	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 118.720%	
36) 1,2-Dichloropropane-d6	6.072	67	51776	60.438	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	= 120.880%#	
41) Toluene-d8	7.317	98	128250	49.904	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 99.800%	
43) trans-1,3-Dichloroprop...	7.628	79	16325	42.689	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 85.380%	
47) 2-Hexanone-d5	8.095	63	24900	79.279	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	= 79.280%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	64621	56.564	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	= 113.120%	
66) 1,2-Dichlorobenzene-d4	11.625	152	45024	60.701	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 121.400%#	
Target Compounds						
					Qvalue	
51) Chlorobenzene	8.886	112	5748	3.106	ug/L	95
65) 1,4-Dichlorobenzene	11.275	146	4801	4.273	ug/L	92
67) 1,2-Dichlorobenzene	11.644	146	7258	6.545	ug/L	90
70) 1,2,4-trichlorobenzene	13.265	180	1978	3.138	ug/L	92
72) 1,2,3-Trichlorobenzene	13.744	180	608	0.946	ug/L #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052622\
Data File : VV025960.D
Acq On : 26 May 2022 15:59
Operator : SY/MD
Sample : VIBLK271
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK271

Quant Time: May 27 01:29:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051022WMA.M
Quant Title : VOC Analysis
QLast Update : Fri May 27 01:28:09 2022
Response via : Initial Calibration

