

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW052622\
 Data File : VW025964.D
 Acq On : 26 May 2022 17:35
 Operator : SY/MD
 Sample : N3039-02
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0JP4

Quant Time: May 27 01:30:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri May 27 01:28:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	334130	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	339774	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	147344	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.310	65	185421	42.998	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.000%
7) Chloroethane-d5	1.571	69	153525	40.541	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	81.080%
11) 1,1-Dichloroethene-d2	2.111	63	262787	34.843	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.680%
21) 2-Butanone-d5	3.908	46	120234	65.052	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	65.050%
24) Chloroform-d	4.349	84	280588	46.585	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.180%
26) 1,2-Dichloroethane-d4	5.034	65	172880	47.492	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.980%
32) Benzene-d6	5.050	84	564823	47.684	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.360%
36) 1,2-Dichloropropane-d6	6.069	67	181694	49.942	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.880%
41) Toluene-d8	7.317	98	475480	43.567	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.140%
43) trans-1,3-Dichloroprop...	7.622	79	66150	40.733	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.460%
47) 2-Hexanone-d5	8.092	63	107405	80.525	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	80.530%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	221088	45.570	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.140%
66) 1,2-Dichlorobenzene-d4	11.622	152	170336	52.701	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052622\
Data File : VV025964.D
Acq On : 26 May 2022 17:35
Operator : SY/MD
Sample : N3039-02
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0JP4

Quant Time: May 27 01:30:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051022WMA.M
Quant Title : VOC Analysis
QLast Update : Fri May 27 01:28:09 2022
Response via : Initial Calibration

