




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052622\
 Data File : VV025958.D
 Acq On : 26 May 2022 15:11
 Operator : SY/MD
 Sample : N3039-01
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0JN7

Manual IntegrationsAPPROVED

Quant Time: May 27 01:29:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri May 27 01:28:09 2022
 Response via : Initial Calibration

Reviewed By :Krupa Patel 05/27/2022
 Supervised By :Mahesh Dadoda 05/27/2022



TIC: VV025958.D\data.ms

(13) Acetone (T)

2.191min (-0.006) 1.81 ug/L

response 2380

Ion	Exp%	Act%
43.00	100.00	100.00
58.00	32.50	23.36
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052622\
 Data File : VV025958.D
 Acq On : 26 May 2022 15:11
 Operator : SY/MD
 Sample : N3039-01
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0JN7

Manual IntegrationsAPPROVED

Quant Time: May 27 01:29:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri May 27 01:28:09 2022
 Response via : Initial Calibration

Reviewed By :Krupa Patel 05/27/2022
 Supervised By :Mahesh Dadoda 05/27/2022



TIC: VV025958.D\data.ms

(13) Acetone (T)

2.204min (+ 0.007) 7.97 ug/L m

response 10453

Ion	Exp%	Act%
43.00	100.00	100.00
58.00	32.50	5.32
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052622\
 Data File : VV025958.D
 Acq On : 26 May 2022 15:11
 Operator : SY/MD
 Sample : N3039-01
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0JN7

Manual IntegrationsAPPROVED

Quant Time: May 27 01:29:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri May 27 01:28:09 2022
 Response via : Initial Calibration

Reviewed By :Krupa Patel 05/27/2022
 Supervised By :Mahesh Dadoda 05/27/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	347422	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	349111	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	161376	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.310	65	177811	39.656	ug/L	0.00
Spiked Amount 50.000	Range 60	- 135	Recovery	=	79.320%	
7) Chloroethane-d5	1.571	69	143722	36.500	ug/L	0.00
Spiked Amount 50.000	Range 70	- 130	Recovery	=	73.000%	
11) 1,1-Dichloroethene-d2	2.111	63	249035	31.756	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	63.520%	
21) 2-Butanone-d5	3.905	46	110994	57.755	ug/L	0.00
Spiked Amount 100.000	Range 40	- 130	Recovery	=	57.750%	
24) Chloroform-d	4.349	84	258817	41.327	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	82.660%	
26) 1,2-Dichloroethane-d4	5.031	65	160096	42.298	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	84.600%	
32) Benzene-d6	5.050	84	530786	43.612	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	87.220%	
36) 1,2-Dichloropropane-d6	6.069	67	168286	45.020	ug/L	0.00
Spiked Amount 50.000	Range 70	- 120	Recovery	=	90.040%	
41) Toluene-d8	7.317	98	453896	40.477	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	80.960%	
43) trans-1,3-Dichloroprop...	7.622	79	62325	37.351	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	74.700%	
47) 2-Hexanone-d5	8.092	63	97053	70.818	ug/L	0.00
Spiked Amount 100.000	Range 45	- 130	Recovery	=	70.820%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	205698	41.264	ug/L	0.00
Spiked Amount 50.000	Range 65	- 120	Recovery	=	82.520%	
66) 1,2-Dichlorobenzene-d4	11.622	152	168976	47.735	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	95.460%	
Target Compounds						
						Qvalue
13) Acetone	2.204	43	10453m	7.971	ug/L	
51) Chlorobenzene	8.883	112	72836	9.020	ug/L	97
64) 1,3-Dichlorobenzene	11.181	146	13322	2.604	ug/L	98
65) 1,4-Dichlorobenzene	11.272	146	50812	9.476	ug/L	98
67) 1,2-Dichlorobenzene	11.641	146	383355	72.435	ug/L	98
70) 1,2,4-trichlorobenzene	13.259	180	74714	24.835	ug/L	98
72) 1,2,3-Trichlorobenzene	13.744	180	44841	14.627	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052622\
Data File : VV025958.D
Acq On : 26 May 2022 15:11
Operator : SY/MD
Sample : N3039-01
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0JN7

Manual IntegrationsAPPROVED

Reviewed By :Krupa Patel 05/27/2022
Supervised By :Mahesh Dadoda 05/27/2022

Quant Time: May 27 01:29:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051022WMA.M
Quant Title : VOC Analysis
QLast Update : Fri May 27 01:28:09 2022
Response via : Initial Calibration

