

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW052622\
 Data File : VW025962.D
 Acq On : 26 May 2022 16:47
 Operator : SY/MD
 Sample : N3039-03
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0JQ0

Quant Time: May 27 01:30:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri May 27 01:28:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	129664	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	134768	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	56734	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.310	65	112580	67.275	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 134.540%		
7) Chloroethane-d5	1.571	69	91686	62.389	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 124.780%		
11) 1,1-Dichloroethene-d2	2.111	63	159061	54.347	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 108.700%		
21) 2-Butanone-d5	3.912	46	76720	106.964	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 106.960%		
24) Chloroform-d	4.352	84	162008	69.313	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 138.620%#		
26) 1,2-Dichloroethane-d4	5.034	65	101615	71.934	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 143.860%#		
32) Benzene-d6	5.053	84	319631	68.032	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 136.060%#		
36) 1,2-Dichloropropane-d6	6.072	67	100077	69.353	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 138.700%#		
41) Toluene-d8	7.317	98	254730	58.846	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 117.700%		
43) trans-1,3-Dichloroprop...	7.625	79	35188	54.627	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 109.260%		
47) 2-Hexanone-d5	8.092	63	54256	102.555	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 102.560%		
56) 1,1,2,2-Tetrachloroeth...	10.214	84	126244	65.604	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 131.200%#		
66) 1,2-Dichlorobenzene-d4	11.622	152	89229	71.699	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 143.400%#		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052622\
Data File : VV025962.D
Acq On : 26 May 2022 16:47
Operator : SY/MD
Sample : N3039-03
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0JQ0

Quant Time: May 27 01:30:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051022WMA.M
Quant Title : VOC Analysis
QLast Update : Fri May 27 01:28:09 2022
Response via : Initial Calibration

