

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW052622\
 Data File : VW025965.D
 Acq On : 26 May 2022 17:59
 Operator : SY/MD
 Sample : N3039-03RE
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0JQ0RE

Quant Time: May 27 01:30:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri May 27 01:28:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	163221	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	167177	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	69342	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.310	65	139623	66.281	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	132.560%
7) Chloroethane-d5	1.571	69	115315	62.336	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	124.680%
11) 1,1-Dichloroethene-d2	2.111	63	195256	52.998	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	106.000%
21) 2-Butanone-d5	3.912	46	94917	105.127	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	105.130%
24) Chloroform-d	4.352	84	203806	69.269	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	138.540%#
26) 1,2-Dichloroethane-d4	5.034	65	124622	70.083	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	140.160%#
32) Benzene-d6	5.050	84	410742	70.477	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	140.960%#
36) 1,2-Dichloropropane-d6	6.069	67	131632	73.537	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	147.080%#
41) Toluene-d8	7.317	98	339218	63.172	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	126.340%#
43) trans-1,3-Dichloroprop...	7.622	79	46224	57.849	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	115.700%
47) 2-Hexanone-d5	8.092	63	75359	114.830	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	114.830%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	163234	68.382	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	136.760%#
66) 1,2-Dichlorobenzene-d4	11.622	152	118960	78.208	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	156.420%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052622\
Data File : VV025965.D
Acq On : 26 May 2022 17:59
Operator : SY/MD
Sample : N3039-03RE
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0JQ0RE

Quant Time: May 27 01:30:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051022WMA.M
Quant Title : VOC Analysis
QLast Update : Fri May 27 01:28:09 2022
Response via : Initial Calibration

