

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW052622\
 Data File : VW025968.D
 Acq On : 26 May 2022 19:10
 Operator : SY/MD
 Sample : N3039-12
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: May 27 01:31:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051022WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri May 27 01:28:09 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Krupa Patel 05/27/2022
 Supervised By :Mahesh Dadoda 05/27/2022

05/27/2022
 Supervised By :Mahesh
 Dadoda

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	204346	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	204812	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	87119	50.000	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.310	65	176072	66.763	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	= 133.520%	
7) Chloroethane-d5	1.571	69	146419	63.220	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	= 126.440%	
11) 1,1-Dichloroethene-d2	2.111	63	249938m	54.187	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 108.380%	
21) 2-Butanone-d5	3.905	46	122502	108.374	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	= 108.370%	
24) Chloroform-d	4.349	84	262444	71.247	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 142.500%#	
26) 1,2-Dichloroethane-d4	5.034	65	165494	74.338	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 148.680%#	
32) Benzene-d6	5.050	84	535942	75.061	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 150.120%#	
36) 1,2-Dichloropropane-d6	6.069	67	171467	78.189	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	= 156.380%#	
41) Toluene-d8	7.317	98	449783	68.370	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 136.740%#	
43) trans-1,3-Dichloroprop...	7.622	79	62112	63.449	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 126.900%#	
47) 2-Hexanone-d5	8.091	63	103172	128.323	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	= 128.320%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	213252	72.919	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	= 145.840%#	
66) 1,2-Dichlorobenzene-d4	11.622	152	156760	82.030	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 164.060%#	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV052622\
Data File : VV025968.D
Acq On : 26 May 2022 19:10
Operator : SY/MD
Sample : N3039-12
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: May 27 01:31:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM051022WMA.M
Quant Title : VOC Analysis
QLast Update : Fri May 27 01:28:09 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 05/27/2022
Supervised By :Mahesh Dadoda 05/27/2022

05/27/2022
Supervised By :Mahesh
Dadoda

05/27/2022

