

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW060822\
 Data File : VW026033.D
 Acq On : 08 Jun 2022 20:13
 Operator : SY/MD
 Sample : N3253-07
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 CODH0

Quant Time: Jun 09 05:24:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR060822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jun 09 05:13:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.625	114	187123	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	182537	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	72336	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.320	65	50037	4.555	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	91.200%	
7) Chloroethane-d5	1.581	69	55267	4.813	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	96.200%	
11) 1,1-Dichloroethene-d2	2.121	63	80081	3.156	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	63.200%	
20) 2-Butanone-d5	3.921	46	66715	41.065	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	82.120%	
24) Chloroform-d	4.362	84	112465	4.831	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	96.600%	
26) 1,2-Dichloroethane-d4	5.047	65	48143	5.070	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	101.400%	
32) Benzene-d6	5.060	84	203575	4.615	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	92.400%	
36) 1,2-Dichloropropane-d6	6.076	67	67090	4.916	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	98.400%	
41) Toluene-d8	7.320	98	156565	4.017	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	80.400%	
43) trans-1,3-Dichloroprop...	7.629	79	16548	4.119	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	82.400%	
46) 2-Hexanone-d5	8.092	63	72801	41.177	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	82.360%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	41825	4.546	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	91.000%	
66) 1,2-Dichlorobenzene-d4	11.625	152	56944	5.145	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	103.000%	
Target Compounds						
19) 1,1-Dichloroethane	3.201	63	3707	0.144	ug/L	96
25) Chloroform	4.391	83	5361	0.202	ug/L	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV060822\
Data File : VV026033.D
Acq On : 08 Jun 2022 20:13
Operator : SY/MD
Sample : N3253-07
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
CODH0

Quant Time: Jun 09 05:24:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR060822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jun 09 05:13:00 2022
Response via : Initial Calibration

