

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW080322\
 Data File : VW027199.D
 Acq On : 03 Aug 2022 18:16
 Operator : SY/MD
 Sample : N4005-03DL 5X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EY146DL

Quant Time: Aug 04 03:05:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Aug 04 02:57:16 2022
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		5.613	114	539986	50.000	ug/L	0.00
28) Chlorobenzene-d5		8.850	117	524158	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		11.249	152	210449	50.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.301	65	223107	63.753	ug/L	0.00
Spiked Amount	50.000	Range 60	- 135	Recovery	=	127.500%	
7) Chloroethane-d5		1.561	69	176204	67.285	ug/L	0.00
Spiked Amount	50.000	Range 70	- 130	Recovery	=	134.560%#	
11) 1,1-Dichloroethene-d2		2.101	63	271312	40.713	ug/L	0.00
Spiked Amount	50.000	Range 60	- 125	Recovery	=	81.420%	
21) 2-Butanone-d5		3.873	46	246877	107.149	ug/L	0.00
Spiked Amount	100.000	Range 40	- 130	Recovery	=	107.150%	
24) Chloroform-d		4.343	84	363277	51.355	ug/L	0.00
Spiked Amount	50.000	Range 70	- 125	Recovery	=	102.700%	
26) 1,2-Dichloroethane-d4		5.027	65	213100	53.096	ug/L	0.00
Spiked Amount	50.000	Range 70	- 125	Recovery	=	106.200%	
32) Benzene-d6		5.047	84	725299	52.293	ug/L	0.00
Spiked Amount	50.000	Range 70	- 125	Recovery	=	104.580%	
36) 1,2-Dichloropropane-d6		6.066	67	237410	55.333	ug/L	0.00
Spiked Amount	50.000	Range 70	- 120	Recovery	=	110.660%	
41) Toluene-d8		7.313	98	612092	48.951	ug/L	0.00
Spiked Amount	50.000	Range 80	- 120	Recovery	=	97.900%	
43) trans-1,3-Dichloroprop...		7.619	79	94475	50.170	ug/L	0.00
Spiked Amount	50.000	Range 60	- 125	Recovery	=	100.340%	
47) 2-Hexanone-d5		8.088	63	181652	109.663	ug/L	0.00
Spiked Amount	100.000	Range 45	- 130	Recovery	=	109.660%	
56) 1,1,2,2-Tetrachloroeth...		10.214	84	335043	52.239	ug/L	0.00
Spiked Amount	50.000	Range 65	- 120	Recovery	=	104.480%	
66) 1,2-Dichlorobenzene-d4		11.622	152	217852	58.902	ug/L	0.00
Spiked Amount	50.000	Range 80	- 120	Recovery	=	117.800%	
Target Compounds							
17) trans-1,2-Dichloroethene		2.754	96	328684	82.468	ug/L	98
20) cis-1,2-Dichloroethene		3.905	96	174359	41.213	ug/L	98
34) Trichloroethene		5.918	95	15728	3.804	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080322\
Data File : VV027199.D
Acq On : 03 Aug 2022 18:16
Operator : SY/MD
Sample : N4005-03DL 5X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EY146DL

Quant Time: Aug 04 03:05:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080122WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Aug 04 02:57:16 2022
Response via : Initial Calibration

