

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW080322\
 Data File : VW027200.D
 Acq On : 03 Aug 2022 18:41
 Operator : SY/MD
 Sample : N4005-05DL 4X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EY157DL

Quant Time: Aug 04 03:05:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Aug 04 02:57:16 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	535303	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	531290	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	213395	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	232061	66.891	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 133.780%		
7) Chloroethane-d5	1.561	69	174823	67.341	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 134.680%#		
11) 1,1-Dichloroethene-d2	2.101	63	274836	41.603	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 83.200%		
21) 2-Butanone-d5	3.873	46	251507	110.114	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 110.110%		
24) Chloroform-d	4.342	84	363189	51.792	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 103.580%		
26) 1,2-Dichloroethane-d4	5.027	65	215490	54.161	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 108.320%		
32) Benzene-d6	5.047	84	736268	52.371	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 104.740%		
36) 1,2-Dichloropropane-d6	6.066	67	237430	54.594	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 109.180%		
41) Toluene-d8	7.313	98	617261	48.702	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 97.400%		
43) trans-1,3-Dichloroprop...	7.619	79	94731	49.631	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 99.260%		
47) 2-Hexanone-d5	8.088	63	182667	108.796	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 108.800%		
56) 1,1,2,2-Tetrachloroeth...	10.214	84	335762	51.648	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 103.300%		
66) 1,2-Dichlorobenzene-d4	11.625	152	222887	59.431	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 118.860%		
Target Compounds						Qvalue
17) trans-1,2-Dichloroethene	2.754	96	295127	74.696	ug/L	98
20) cis-1,2-Dichloroethene	3.905	96	209635	49.984	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080322\
Data File : VV027200.D
Acq On : 03 Aug 2022 18:41
Operator : SY/MD
Sample : N4005-05DL 4X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EY157DL

Quant Time: Aug 04 03:05:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080122WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Aug 04 02:57:16 2022
Response via : Initial Calibration

