




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080322\
 Data File : VV027196.D
 Acq On : 03 Aug 2022 17:02
 Operator : SY/MD
 Sample : N4003-01DL 4X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F5E13DL

Manual IntegrationsAPPROVED

Quant Time: Aug 04 03:05:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Aug 04 02:57:16 2022
 Response via : Initial Calibration

Reviewed By :Krupa Patel 08/04/2022
 Supervised By :Mahesh Dadoda 08/05/2022



TIC: VV027196.D\data.ms

(11) 1,1-Dichloroethene-d2 (S)

2.204min (+ 0.103) 2.23 ug/L

response 18174

Ion	Exp%	Act%
63.00	100.00	100.00
98.00	79.90	71.20
65.00	23.50	28.06
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080322\
Data File : VV027196.D
Acq On : 03 Aug 2022 17:02
Operator : SY/MD
Sample : N4003-01DL 4X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F5E13DL

Manual IntegrationsAPPROVED

Reviewed By :Krupa Patel 08/04/2022
Supervised By :Mahesh Dadoda 08/05/2022

Quant Time: Aug 04 03:05:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080122WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Aug 04 02:57:16 2022
Response via : Initial Calibration



TIC: VV027196.D\data.ms

(11) 1,1-Dichloroethene-d2 (S)

2.101min (-0.000) 38.65 ug/L m

response 314481

Ion	Exp%	Act%
63.00	100.00	100.00
98.00	79.90	4.11#
65.00	23.50	1.62#
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080322\
 Data File : VV027196.D
 Acq On : 03 Aug 2022 17:02
 Operator : SY/MD
 Sample : N4003-01DL 4X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F5E13DL

Manual IntegrationsAPPROVED

Quant Time: Aug 04 03:05:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080122WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Aug 04 02:57:16 2022
 Response via : Initial Calibration

Reviewed By :Krupa Patel 08/04/2022
 Supervised By :Mahesh Dadoda 08/05/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	659238	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	679788	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	284392	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	259146	60.655	ug/L	0.00
Spiked Amount 50.000	Range 60	- 135	Recovery	=	121.320%	
7) Chloroethane-d5	1.558	69	199356	62.355	ug/L	0.00
Spiked Amount 50.000	Range 70	- 130	Recovery	=	124.700%	
11) 1,1-Dichloroethene-d2	2.101	63	314481m	38.654	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	77.300%	
21) 2-Butanone-d5	3.873	46	275317	97.877	ug/L	0.00
Spiked Amount 100.000	Range 40	- 130	Recovery	=	97.880%	
24) Chloroform-d	4.343	84	411575	47.658	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	95.320%	
26) 1,2-Dichloroethane-d4	5.027	65	244233	49.845	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	99.700%	
32) Benzene-d6	5.043	84	839539	46.672	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	93.340%	
36) 1,2-Dichloropropane-d6	6.066	67	278666	50.079	ug/L	0.00
Spiked Amount 50.000	Range 70	- 120	Recovery	=	100.160%	
41) Toluene-d8	7.313	98	721512	44.492	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	88.980%	
43) trans-1,3-Dichloroprop...	7.619	79	111723	45.747	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	91.500%	
47) 2-Hexanone-d5	8.088	63	216633	100.840	ug/L	0.00
Spiked Amount 100.000	Range 45	- 130	Recovery	=	100.840%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	382659	46.004	ug/L	0.00
Spiked Amount 50.000	Range 65	- 120	Recovery	=	92.000%	
66) 1,2-Dichlorobenzene-d4	11.625	152	261563	52.333	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	104.660%	
Target Compounds						
					Qvalue	
5) Vinyl chloride	1.307	62	325000	52.198	ug/L	99
17) trans-1,2-Dichloroethene	2.757	96	13086	2.689	ug/L	97
20) cis-1,2-Dichloroethene	3.908	96	4956	0.960	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV080322\
Data File : VV027196.D
Acq On : 03 Aug 2022 17:02
Operator : SY/MD
Sample : N4003-01DL 4X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 3

Instrument :
MSVOA_V
ClientSampleId :
F5E13DL

Manual IntegrationsAPPROVED

Quant Time: Aug 04 03:05:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM080122WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Aug 04 02:57:16 2022
Response via : Initial Calibration

Reviewed By :Krupa Patel 08/04/2022
Supervised By :Mahesh Dadoda 08/05/2022

