

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV082125\
 Data File : VV038983.D
 Acq On : 21 Aug 2025 16:12
 Operator : SY/MD
 Sample : VV0821MBL01
 Misc : 5.00g/10.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VV0821MBL01

Quant Time: Aug 22 04:32:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 04:15:04 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.597	168	427239	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.532	114	741015	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.773	117	612117	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.175	152	211578	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.937	65	309584	51.641	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	103.280%
35) Dibromofluoromethane	4.500	113	270239	48.628	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	97.260%
50) Toluene-d8	7.236	98	842767	43.490	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	86.980%#
62) 4-Bromofluorobenzene	10.001	95	286884	40.468	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	80.940%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV082125\
Data File : VV038983.D
Acq On : 21 Aug 2025 16:12
Operator : SY/MD
Sample : VV0821MBL01
Misc : 5.00g/10.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VV0821MBL01

Quant Time: Aug 22 04:32:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 04:15:04 2025
Response via : Initial Calibration

