

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV082125\
 Data File : VV038984.D
 Acq On : 21 Aug 2025 16:34
 Operator : SY/MD
 Sample : VV0821WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VV0821WBL01

Quant Time: Aug 22 04:32:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 04:15:04 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.600	168	502511	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.532	114	836910	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	730493	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.175	152	318475	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.940	65	339699	48.177	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	96.360%
35) Dibromofluoromethane	4.503	113	274834	43.789	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	87.580%
50) Toluene-d8	7.236	98	950290	43.420	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	86.840%#
62) 4-Bromofluorobenzene	10.001	95	338753	42.310	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	84.620%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV082125\
Data File : VV038984.D
Acq On : 21 Aug 2025 16:34
Operator : SY/MD
Sample : VV0821WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VV0821WBL01

Quant Time: Aug 22 04:32:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 04:15:04 2025
Response via : Initial Calibration

