

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV082125\
 Data File : VV038988.D
 Acq On : 21 Aug 2025 18:14
 Operator : SY/MD
 Sample : Q2929-02
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DA5-SW3

Quant Time: Aug 22 04:35:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 04:15:04 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.600	168	567969	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.532	114	999329	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	1081937	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.175	152	465817	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.941	65	394780	49.536	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	99.080%
35) Dibromofluoromethane	4.500	113	322374	43.015	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	86.040%
50) Toluene-d8	7.236	98	1193111	45.654	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	91.300%
62) 4-Bromofluorobenzene	10.002	95	485086	50.740	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	101.480%
Target Compounds						
16) Acetone	2.124	43	33120	10.354	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV082125\
Data File : VV038988.D
Acq On : 21 Aug 2025 18:14
Operator : SY/MD
Sample : Q2929-02
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DA5-SW3

Quant Time: Aug 22 04:35:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 04:15:04 2025
Response via : Initial Calibration

