

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV082125\
 Data File : VV038990.D
 Acq On : 21 Aug 2025 18:57
 Operator : SY/MD
 Sample : Q2929-04
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DA6-SW5

Quant Time: Aug 22 04:36:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 04:15:04 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.600	168	627795	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.532	114	1158287	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	1037879	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.175	152	449449	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.940	65	477148	54.166	ug/l	0.00
Spiked Amount	50.000	Range 81 - 118	Recovery	=	108.340%	
35) Dibromofluoromethane	4.503	113	404529	46.570	ug/l	0.00
Spiked Amount	50.000	Range 80 - 119	Recovery	=	93.140%	
50) Toluene-d8	7.236	98	1352656	44.656	ug/l	0.00
Spiked Amount	50.000	Range 89 - 112	Recovery	=	89.320%	
62) 4-Bromofluorobenzene	10.001	95	431386	38.930	ug/l	0.00
Spiked Amount	50.000	Range 85 - 114	Recovery	=	77.860%#	
Target Compounds						
16) Acetone	2.124	43	22733	6.430	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV082125\
Data File : VV038990.D
Acq On : 21 Aug 2025 18:57
Operator : SY/MD
Sample : Q2929-04
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DA6-SW5

Quant Time: Aug 22 04:36:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 04:15:04 2025
Response via : Initial Calibration

