

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV082125\
 Data File : VV038991.D
 Acq On : 21 Aug 2025 19:18
 Operator : SY/MD
 Sample : Q2930-01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DA1-1

Quant Time: Aug 22 04:36:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 04:15:04 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.600	168	584299	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.532	114	1018996	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	897248	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.175	152	391469	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.941	65	416123	50.755	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	101.500%
35) Dibromofluoromethane	4.500	113	352048	46.068	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	92.140%
50) Toluene-d8	7.236	98	1183190	44.401	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	88.800%#
62) 4-Bromofluorobenzene	10.001	95	373770	38.341	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	76.680%#
Target Compounds						
16) Acetone	2.127	43	13859	4.212	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV082125\
Data File : VV038991.D
Acq On : 21 Aug 2025 19:18
Operator : SY/MD
Sample : Q2930-01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DA1-1

Quant Time: Aug 22 04:36:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 04:15:04 2025
Response via : Initial Calibration

