

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV082125\
 Data File : VV038999.D
 Acq On : 21 Aug 2025 22:11
 Operator : SY/MD
 Sample : Q2924-06
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 MW-2A-082025

Quant Time: Aug 22 04:40:02 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V082125W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 22 04:15:04 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	4.600	168	650572	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	5.532	114	1153105	50.000	ug/l	0.00
63) Chlorobenzene-d5	8.776	117	1016294	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.175	152	448609	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.940	65	449834	49.277	ug/l	0.00
Spiked Amount	50.000	Range	81 - 118	Recovery	=	98.560%
35) Dibromofluoromethane	4.503	113	385583	44.588	ug/l	0.00
Spiked Amount	50.000	Range	80 - 119	Recovery	=	89.180%
50) Toluene-d8	7.236	98	1326138	43.977	ug/l	0.00
Spiked Amount	50.000	Range	89 - 112	Recovery	=	87.960%#
62) 4-Bromofluorobenzene	10.001	95	462072	41.887	ug/l	0.00
Spiked Amount	50.000	Range	85 - 114	Recovery	=	83.780%#
Target Compounds						
					Qvalue	
16) Acetone	2.117	43	711942	194.315	ug/l	98
30) Chloroform	4.291	83	35087	2.141	ug/l	92
40) Benzene	5.008	78	1078206	30.653	ug/l	99
44) Trichloroethene	5.857	130	4604	0.483	ug/l	61
64) Tetrachloroethene	7.902	164	40507	5.357	ug/l	94
67) Ethyl Benzene	8.953	91	17096	0.457	ug/l	93
73) Isopropylbenzene	9.866	105	7927	0.263	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV082125\
Data File : VV038999.D
Acq On : 21 Aug 2025 22:11
Operator : SY/MD
Sample : Q2924-06
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
MW-2A-082025

Quant Time: Aug 22 04:40:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\82V082125W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 22 04:15:04 2025
Response via : Initial Calibration

