




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090122\
 Data File : VV027709.D
 Acq On : 02 Sep 2022 09:09
 Operator : SY/MD
 Sample : N4404-13
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 39 Sample Multiplier: 1

Quant Time: Sep 03 04:15:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR083122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Sep 02 03:03:49 2022
 Response via : Initial Calibration



TIC: VV027709.D\data.ms

(52) Ethylbenzene (T)

9.140min (+ 0.129) 4.14 ug/L

response 207451

Ion	Exp%	Act%
91.15	100.00	100.00
106.15	31.30	50.04#
0.00	0.00	0.00
0.00	0.00	0.00