

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090122\
 Data File : VV027674.D
 Acq On : 01 Sep 2022 18:08
 Operator : SY/MD
 Sample : VV0901WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK222

Quant Time: Sep 02 00:18:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR083122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Sep 02 00:17:48 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.596	114	140976	5.000	ug/L	-0.01
28) Chlorobenzene-d5	8.863	117	118972	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	49572	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.297	65	61294	4.002	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.000%
7) Chloroethane-d5	1.558	69	48369	3.775	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	75.400%
11) 1,1-Dichloroethene-d2	2.098	63	91076	3.038	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.800%
20) 2-Butanone-d5	3.957	46	126277	48.767	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.540%
24) Chloroform-d	4.352	84	111806	3.636	ug/L	0.01
Spiked Amount	5.000	Range	70 - 125	Recovery	=	72.800%
26) 1,2-Dichloroethane-d4	5.034	65	61201	4.444	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.800%
32) Benzene-d6	5.037	84	215503	4.422	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.400%
36) 1,2-Dichloropropane-d6	6.063	67	67980	4.510	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.200%
41) Toluene-d8	7.333	98	177986	4.217	ug/L	0.02
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.400%
43) trans-1,3-Dichloroprop...	7.641	79	19889	4.283	ug/L	0.01
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.600%
46) 2-Hexanone-d5	8.120	63	65414	37.459	ug/L	0.01
Spiked Amount	50.000	Range	45 - 130	Recovery	=	74.920%
56) 1,1,2,2-Tetrachloroeth...	10.220	84	39975	4.121	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	82.400%
66) 1,2-Dichlorobenzene-d4	11.625	152	52510	4.731	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090122\
Data File : VV027674.D
Acq On : 01 Sep 2022 18:08
Operator : SY/MD
Sample : VV0901WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK222

Quant Time: Sep 02 00:18:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR083122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Sep 02 00:17:48 2022
Response via : Initial Calibration

