

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090122\
 Data File : VV027701.D
 Acq On : 02 Sep 2022 05:54
 Operator : SY/MD
 Sample : N4402-15
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBVD4

Quant Time: Sep 03 04:13:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR083122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Sep 02 03:03:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	133711	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	128844	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	56485	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	58312	4.014	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.200%
7) Chloroethane-d5	1.568	69	53090	4.368	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.400%
11) 1,1-Dichloroethene-d2	2.108	63	91738	3.227	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	64.600%
20) 2-Butanone-d5	3.889	46	67477	27.475	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	54.940%
24) Chloroform-d	4.346	84	123882	4.248	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.000%
26) 1,2-Dichloroethane-d4	5.031	65	59444	4.551	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.000%
32) Benzene-d6	5.047	84	234558	4.444	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.800%
36) 1,2-Dichloropropane-d6	6.066	67	70101	4.294	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.800%
41) Toluene-d8	7.314	98	191130	4.182	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.600%
43) trans-1,3-Dichloroprop...	7.625	79	19739	3.925	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.600%
46) 2-Hexanone-d5	8.088	63	91173	48.210	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.420%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	43130	4.106	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	82.200%
66) 1,2-Dichlorobenzene-d4	11.622	152	64573	5.106	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.200%
Target Compounds						Qvalue
22) cis-1,2-Dichloroethene	3.909	96	130179	10.310	ug/L	91
34) Trichloroethene	5.915	95	39970	3.236	ug/L	94
47) Tetrachloroethene	7.973	164	793303	79.202	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090122\
Data File : VV027701.D
Acq On : 02 Sep 2022 05:54
Operator : SY/MD
Sample : N4402-15
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBVD4

Quant Time: Sep 03 04:13:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR083122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Sep 02 03:03:49 2022
Response via : Initial Calibration

