

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV090122\  
 Data File : VV027713.D  
 Acq On : 02 Sep 2022 10:46  
 Operator : SY/MD  
 Sample : N4402-05 100X  
 Misc : 25.0mL/MSVOA\_V/WATER  
 ALS Vial : 43 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 DBVB0

Quant Time: Sep 03 04:16:38 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR083122WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Fri Sep 02 03:03:49 2022  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-------------------------------|--------|-------|----------|----------|-------|----------|
| -----                         |        |       |          |          |       |          |
| Internal Standards            |        |       |          |          |       |          |
| 1) 1,4-Difluorobenzene        | 5.616  | 114   | 121919   | 5.000    | ug/L  | 0.00     |
| 28) Chlorobenzene-d5          | 8.850  | 117   | 117517   | 5.000    | ug/L  | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.249 | 152   | 50301    | 5.000    | ug/L  | 0.00     |
|                               |        |       |          |          |       |          |
| System Monitoring Compounds   |        |       |          |          |       |          |
| 4) Vinyl Chloride-d3          | 0.000  | 65    | 0d       | 0.000    | ug/L  |          |
| Spiked Amount                 | 5.000  | Range | 40 - 130 | Recovery | =     | 0.000%#  |
| 7) Chloroethane-d5            | 1.568  | 69    | 48666    | 4.392    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 130 | Recovery | =     | 87.800%  |
| 11) 1,1-Dichloroethene-d2     | 2.105  | 63    | 297      | 0.011    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 125 | Recovery | =     | 0.200%#  |
| 20) 2-Butanone-d5             | 3.899  | 46    | 64293    | 28.710   | ug/L  | 0.01     |
| Spiked Amount                 | 50.000 | Range | 40 - 130 | Recovery | =     | 57.420%  |
| 24) Chloroform-d              | 4.346  | 84    | 118851   | 4.470    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 89.400%  |
| 26) 1,2-Dichloroethane-d4     | 5.031  | 65    | 54735    | 4.595    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 92.000%  |
| 32) Benzene-d6                | 5.047  | 84    | 215834   | 4.484    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 89.600%  |
| 36) 1,2-Dichloropropane-d6    | 6.066  | 67    | 65549    | 4.402    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 140 | Recovery | =     | 88.000%  |
| 41) Toluene-d8                | 7.313  | 98    | 166830   | 4.002    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 80.000%  |
| 43) trans-1,3-Dichloroprop... | 0.000  | 79    | 0        | 0.000    | ug/L  |          |
| Spiked Amount                 | 5.000  | Range | 55 - 130 | Recovery | =     | 0.000%#  |
| 46) 2-Hexanone-d5             | 8.088  | 63    | 74286    | 43.066   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000 | Range | 45 - 130 | Recovery | =     | 86.140%  |
| 56) 1,1,2,2-Tetrachloroeth... | 10.214 | 84    | 40616    | 4.239    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 120 | Recovery | =     | 84.800%  |
| 66) 1,2-Dichlorobenzene-d4    | 11.622 | 152   | 56888    | 5.052    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 80 - 120 | Recovery | =     | 101.000% |
|                               |        |       |          |          |       |          |
| Target Compounds              |        |       |          |          |       | Qvalue   |
| 16) Methylene chloride        | 2.507  | 84    | 11721    | 0.913    | ug/L  | 96       |
| -----                         |        |       |          |          |       |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV090122\  
Data File : VV027713.D  
Acq On : 02 Sep 2022 10:46  
Operator : SY/MD  
Sample : N4402-05 100X  
Misc : 25.0mL/MSVOA\_V/WATER  
ALS Vial : 43 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
DBVB0

Quant Time: Sep 03 04:16:38 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR083122WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Fri Sep 02 03:03:49 2022  
Response via : Initial Calibration

