

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090122\
 Data File : VV027718.D
 Acq On : 02 Sep 2022 12:47
 Operator : SY/MD
 Sample : N4400-17
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Sep 03 04:17:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR083122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Sep 02 03:03:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	125408	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	123202	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	55098	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	57586	4.227	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	84.600%	
7) Chloroethane-d5	1.568	69	48000	4.211	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	84.200%	
11) 1,1-Dichloroethene-d2	2.108	63	90660	3.400	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	68.000%	
20) 2-Butanone-d5	3.896	46	50500	21.924	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	43.840%	
24) Chloroform-d	4.346	84	110357	4.035	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	80.600%	
26) 1,2-Dichloroethane-d4	5.031	65	51842	4.231	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	84.600%	
32) Benzene-d6	5.047	84	209869	4.159	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	83.200%	
36) 1,2-Dichloropropane-d6	6.066	67	60936	3.904	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	78.000%	
41) Toluene-d8	7.313	98	168899	3.865	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	77.200%	
43) trans-1,3-Dichloroprop...	7.625	79	19389	4.032	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	80.600%	
46) 2-Hexanone-d5	8.092	63	65969	36.480	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	72.960%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	35380	3.522	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	70.400%	
66) 1,2-Dichlorobenzene-d4	11.622	152	55292	4.482	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	89.600%	
Target Compounds						Qvalue
16) Methylene chloride	2.503	84	5035	0.381	ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090122\
Data File : VV027718.D
Acq On : 02 Sep 2022 12:47
Operator : SY/MD
Sample : N4400-17
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Sep 03 04:17:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR083122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Sep 02 03:03:49 2022
Response via : Initial Calibration

