

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090122\
 Data File : VV027721.D
 Acq On : 02 Sep 2022 14:31
 Operator : SY/MD
 Sample : VV0901WBL03
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK224

Quant Time: Sep 03 05:32:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR083122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Sep 03 05:30:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	121204	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	120115	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	53004	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	62860	4.774	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	95.400%
7) Chloroethane-d5	1.568	69	50202	4.557	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.200%
11) 1,1-Dichloroethene-d2	2.105	63	95902	3.721	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.400%
20) 2-Butanone-d5	3.899	46	60371	27.118	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	54.240%
24) Chloroform-d	4.342	84	118622	4.488	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.800%
26) 1,2-Dichloroethane-d4	5.031	65	56448	4.767	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.400%
32) Benzene-d6	5.047	84	224836	4.570	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.400%
36) 1,2-Dichloropropane-d6	6.066	67	65642	4.313	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	86.200%
41) Toluene-d8	7.313	98	187440	4.399	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.000%
43) trans-1,3-Dichloroprop...	7.622	79	21572	4.601	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.000%
46) 2-Hexanone-d5	8.091	63	74482	42.246	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.500%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	40430	4.128	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	82.600%
66) 1,2-Dichlorobenzene-d4	11.622	152	61449	5.178	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.600%
Target Compounds						Qvalue
16) Methylene chloride	2.506	84	10703	0.839	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090122\
Data File : VV027721.D
Acq On : 02 Sep 2022 14:31
Operator : SY/MD
Sample : VV0901WBL03
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK224

Quant Time: Sep 03 05:32:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR083122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Sep 03 05:30:57 2022
Response via : Initial Calibration

