

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090122\
 Data File : VV027722.D
 Acq On : 02 Sep 2022 14:58
 Operator : SY/MD
 Sample : N4404-15
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBVD0

Quant Time: Sep 03 05:33:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR083122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Sep 03 05:30:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	121846	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	122766	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	56654	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	52264	3.948	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.000%
7) Chloroethane-d5	1.568	69	49092	4.433	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	88.600%
11) 1,1-Dichloroethene-d2	2.105	63	84318	3.255	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.000%
20) 2-Butanone-d5	3.895	46	61323	27.401	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	54.800%
24) Chloroform-d	4.342	84	115403	4.343	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.800%
26) 1,2-Dichloroethane-d4	5.030	65	54996	4.620	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.400%
32) Benzene-d6	5.043	84	222287	4.420	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.400%
36) 1,2-Dichloropropane-d6	6.066	67	64960	4.176	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	83.600%
41) Toluene-d8	7.313	98	173446	3.983	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.600%
43) trans-1,3-Dichloroprop...	7.622	79	19070	3.980	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.600%
46) 2-Hexanone-d5	8.088	63	79335	44.027	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.060%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	42535	4.249	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	85.000%
66) 1,2-Dichlorobenzene-d4	11.622	152	63893	5.037	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.800%
Target Compounds						
5) Vinyl chloride	1.310	62	7426	0.491	ug/L	# 82
16) Methylene chloride	2.506	84	6324	0.493	ug/L	87
22) cis-1,2-Dichloroethene	3.908	96	29261	2.543	ug/L	88
34) Trichloroethene	5.918	95	5027	0.427	ug/L	95
47) Tetrachloroethene	7.973	164	68699	7.198	ug/L	97
53) m,p-Xylene	9.140	106	100091	5.375	ug/L	94
54) o-Xylene	9.545	106	1196	0.067	ug/L	79
63) 1,2,4-Trimethylbenzene	10.918	105	4104	0.122	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090122\
Data File : VV027722.D
Acq On : 02 Sep 2022 14:58
Operator : SY/MD
Sample : N4404-15
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBVD0

Quant Time: Sep 03 05:33:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR083122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Sep 03 05:30:57 2022
Response via : Initial Calibration

