

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090122\
 Data File : VV027724.D
 Acq On : 02 Sep 2022 15:47
 Operator : SY/MD
 Sample : N4404-18
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBVD9

Quant Time: Sep 03 05:33:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR083122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Sep 03 05:30:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	121945	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	120379	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	55363	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	58488	4.415	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.200%
7) Chloroethane-d5	1.568	69	53972	4.869	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.400%
11) 1,1-Dichloroethene-d2	2.104	63	91608	3.533	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.600%
20) 2-Butanone-d5	3.905	46	59431	26.534	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	53.060%
24) Chloroform-d	4.346	84	119524	4.494	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.800%
26) 1,2-Dichloroethane-d4	5.030	65	57482	4.825	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.400%
32) Benzene-d6	5.047	84	220482	4.471	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.400%
36) 1,2-Dichloropropane-d6	6.066	67	66113	4.335	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	86.600%
41) Toluene-d8	7.313	98	178086	4.170	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.400%
43) trans-1,3-Dichloroprop...	7.622	79	19021	4.048	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.000%
46) 2-Hexanone-d5	8.091	63	80052	45.306	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.620%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	41551	4.233	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	84.600%
66) 1,2-Dichlorobenzene-d4	11.622	152	65181	5.259	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.200%
Target Compounds						
5) Vinyl chloride	1.310	62	17653	1.165	ug/L	89
22) cis-1,2-Dichloroethene	3.908	96	24949	2.167	ug/L	93
47) Tetrachloroethene	7.976	164	775	0.083	ug/L	93
53) m,p-Xylene	9.149	106	3120	0.171	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090122\
Data File : VV027724.D
Acq On : 02 Sep 2022 15:47
Operator : SY/MD
Sample : N4404-18
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBVD9

Quant Time: Sep 03 05:33:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR083122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Sep 03 05:30:57 2022
Response via : Initial Calibration

