

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090122\
 Data File : VV027732.D
 Acq On : 02 Sep 2022 19:02
 Operator : SY/MD
 Sample : N4418-11
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DBVE7

Quant Time: Sep 03 05:35:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR083122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Sep 03 05:30:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	118847	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	114230	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	51116	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	59133	4.580	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	91.600%	
7) Chloroethane-d5	1.568	69	55125	5.103	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	102.000%	
11) 1,1-Dichloroethene-d2	2.105	63	94533	3.741	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	74.800%	
20) 2-Butanone-d5	3.886	46	76005	34.818	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	69.640%	
24) Chloroform-d	4.346	84	121957	4.705	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	94.200%	
26) 1,2-Dichloroethane-d4	5.031	65	58102	5.004	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	100.000%	
32) Benzene-d6	5.047	84	222028	4.745	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	95.000%	
36) 1,2-Dichloropropane-d6	6.066	67	67176	4.641	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	92.800%	
41) Toluene-d8	7.313	98	175191	4.324	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	86.400%	
43) trans-1,3-Dichloroprop...	7.625	79	20774	4.659	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	93.200%	
46) 2-Hexanone-d5	8.088	63	89662	53.476	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	106.960%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	41627	4.470	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	89.400%	
66) 1,2-Dichlorobenzene-d4	11.622	152	61412	5.366	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	107.400%	
Target Compounds						
2) Dichlorodifluoromethane	1.130	85	1125	0.069	ug/L #	82
5) Vinyl chloride	1.310	62	48904	3.313	ug/L	95
9) Trichlorofluoromethane	1.751	101	5035	0.254	ug/L	94
12) 1,1-Dichloroethene	2.114	96	2925	0.295	ug/L #	1
18) trans-1,2-Dichloroethene	2.761	96	8784	0.781	ug/L	88
22) cis-1,2-Dichloroethene	3.905	96	521715	46.488	ug/L	91
34) Trichloroethene	5.908	95	417949	38.161	ug/L	98
47) Tetrachloroethene	7.973	164	417544	47.020	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV090122\
Data File : VV027732.D
Acq On : 02 Sep 2022 19:02
Operator : SY/MD
Sample : N4418-11
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DBVE7

Quant Time: Sep 03 05:35:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR083122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Sep 03 05:30:57 2022
Response via : Initial Calibration

