

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092019\
 Data File : VV012921.D
 Acq On : 20 Sep 2019 13:05
 Operator : SY/MD
 Sample : VSTDCCC0.5EC
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.5251

Quant Time: Sep 20 23:44:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR070819SIM.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Sep 20 23:41:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	10584	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.89	117	9461	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.30	152	3858	0.50	ug/L	0.00

System Monitoring Compounds

2) Vinyl Chloride-d3	1.32	65	2168	0.19	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	38.00%#
4) 1,2-Dichloroethane-d4	5.09	65	4434	0.51	ug/L	0.02
Spiked Amount	0.500	Range	70 - 130	Recovery	=	102.00%
7) 1,2-Dichloropropane-d6	6.12	67	4176	0.46	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	92.00%
8) Toluene-d8	7.35	98	10989	0.46	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	92.00%
10) 1,1,2,2-Tetrachloroethane-	10.25	84	3607	0.63	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	126.00%#

Target Compounds

					Qvalue
3) Vinyl chloride	1.32	62	2830	0.258	ug/L 97
6) Trichloroethene	5.95	95	3147	0.361	ug/L 97
9) 1,2-Dibromoethane	8.40	107	2348	0.498	ug/L 99
12) 1,2,3-Trichloropropane	10.30	75	2759	0.657	ug/L 97
13) 1,2-Dibromo-3-chloropropan	12.46	75	479	0.621	ug/L 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092019\
Data File : VV012921.D
Acq On : 20 Sep 2019 13:05
Operator : SY/MD
Sample : VSTDCCC0.5EC
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5251

Quant Time: Sep 20 23:44:23 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR070819SIM.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Sep 20 23:41:25 2019
Response via : Initial Calibration

