

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV092019\
 Data File : VV012917.D
 Acq On : 20 Sep 2019 10:52
 Operator : SY/MD
 Sample : VV0920WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK111

Quant Time: Sep 20 23:47:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR070819SIM.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Sep 20 23:41:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	11890	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.89	117	10834	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.30	152	4652	0.50	ug/L	0.00

System Monitoring Compounds

2) Vinyl Chloride-d3	1.32	65	2375	0.18	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	36.00%#
4) 1,2-Dichloroethane-d4	5.09	65	3455	0.35	ug/L	0.02
Spiked Amount	0.500	Range	70 - 130	Recovery	=	70.00%
7) 1,2-Dichloropropane-d6	6.12	67	4540	0.43	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	86.00%
8) Toluene-d8	7.35	98	8808	0.32	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	64.00%#
10) 1,1,2,2-Tetrachloroethane-	10.25	84	3738	0.57	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	114.00%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_V\DATA\VV092019\
Data File : VV012917.D
Acq On : 20 Sep 2019 10:52
Operator : SY/MD
Sample : VV0920WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK111

Quant Time: Sep 20 23:47:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR070819SIM.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Sep 20 23:41:25 2019
Response via : Initial Calibration

