

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092822\
 Data File : VV028160.D
 Acq On : 28 Sep 2022 15:38
 Operator : SY/MD
 Sample : N4871-03
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F5H23

Quant Time: Sep 29 01:28:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092222WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Sep 29 01:25:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	151505	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	158493	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	72372	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	47730	3.039	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	60.800%	
7) Chloroethane-d5	1.565	69	46991	3.763	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	75.200%	
11) 1,1-Dichloroethene-d2	2.101	63	65887	2.419	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	48.400%#	
20) 2-Butanone-d5	3.889	46	145568	54.896	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	109.800%	
24) Chloroform-d	4.343	84	106975	4.268	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	85.400%	
26) 1,2-Dichloroethane-d4	5.027	65	52234	4.380	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	87.600%	
32) Benzene-d6	5.043	84	185605	3.571	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	71.400%	
36) 1,2-Dichloropropane-d6	6.066	67	57902	3.574	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	71.400%	
41) Toluene-d8	7.313	98	145148	3.086	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	61.800%#	
43) trans-1,3-Dichloroprop...	7.622	79	20498	3.503	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	70.000%	
46) 2-Hexanone-d5	8.085	63	107160	43.873	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	87.740%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	51348	4.637	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	92.800%	
66) 1,2-Dichlorobenzene-d4	11.622	152	63961	4.506	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	90.200%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092822\
Data File : VV028160.D
Acq On : 28 Sep 2022 15:38
Operator : SY/MD
Sample : N4871-03
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F5H23

Quant Time: Sep 29 01:28:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092222WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Sep 29 01:25:35 2022
Response via : Initial Calibration

