

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW092822\
 Data File : VW028164.D
 Acq On : 28 Sep 2022 17:16
 Operator : SY/MD
 Sample : N4871-07
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F5H57

Quant Time: Sep 29 01:29:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092222WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Sep 29 01:25:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	150798	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	141312	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	70052	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.301	65	49784	3.184	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	63.600%
7) Chloroethane-d5	1.561	69	49741	4.002	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	80.000%
11) 1,1-Dichloroethene-d2	2.101	63	67706	2.497	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	50.000%#
20) 2-Butanone-d5	3.889	46	160258	60.719	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	121.440%
24) Chloroform-d	4.342	84	118626	4.755	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.200%
26) 1,2-Dichloroethane-d4	5.027	65	55560	4.680	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.600%
32) Benzene-d6	5.043	84	201021	4.338	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.800%
36) 1,2-Dichloropropane-d6	6.066	67	64082	4.436	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.800%
41) Toluene-d8	7.313	98	146187	3.486	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	69.800%#
43) trans-1,3-Dichloroprop...	7.622	79	20508	3.930	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.600%
46) 2-Hexanone-d5	8.085	63	109055	50.077	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.160%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	47955	4.857	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.200%
66) 1,2-Dichlorobenzene-d4	11.622	152	65982	4.802	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV092822\
Data File : VV028164.D
Acq On : 28 Sep 2022 17:16
Operator : SY/MD
Sample : N4871-07
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F5H57

Quant Time: Sep 29 01:29:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092222WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Sep 29 01:25:35 2022
Response via : Initial Calibration

