

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100822\
 Data File : VV028442.D
 Acq On : 08 Oct 2022 17:25
 Operator : SY/MD
 Sample : VV1008WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK271

Quant Time: Oct 09 05:52:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM100722WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Oct 08 02:10:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	345243	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	316403	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.242	152	160587	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	91845	34.477	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	68.960%
7) Chloroethane-d5	1.564	69	72786	34.459	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	68.920%#
11) 1,1-Dichloroethene-d2	2.101	63	125686	27.722	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	55.440%#
21) 2-Butanone-d5	3.895	46	147763	63.363	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	63.360%
24) Chloroform-d	4.346	84	210419	43.068	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.140%
26) 1,2-Dichloroethane-d4	5.027	65	142768	45.310	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.620%
32) Benzene-d6	5.047	84	437894	43.191	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.380%
36) 1,2-Dichloropropane-d6	6.066	67	138406	41.870	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	83.740%
41) Toluene-d8	7.310	98	401027	44.179	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.360%
43) trans-1,3-Dichloroprop...	7.619	79	71761	42.591	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.180%
47) 2-Hexanone-d5	8.085	63	141052	78.441	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	78.440%
56) 1,1,2,2-Tetrachloroeth...	10.210	84	181132	40.043	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	80.080%
66) 1,2-Dichlorobenzene-d4	11.615	152	146006	43.879	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.760%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100822\
Data File : VV028442.D
Acq On : 08 Oct 2022 17:25
Operator : SY/MD
Sample : VV1008WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK271

Quant Time: Oct 09 05:52:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM100722WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Oct 08 02:10:35 2022
Response via : Initial Calibration

