

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100822\
 Data File : VV028443.D
 Acq On : 08 Oct 2022 17:49
 Operator : SY/MD
 Sample : VV1008WBL02
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK272

Quant Time: Oct 09 05:52:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM100722WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Oct 08 02:10:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	330950	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	300401	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	149612	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	127226	49.822	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	99.640%
7) Chloroethane-d5	1.564	69	91384	45.132	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	90.260%
11) 1,1-Dichloroethene-d2	2.105	63	197455	45.433	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.860%
21) 2-Butanone-d5	3.902	46	163262	73.033	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	73.030%
24) Chloroform-d	4.342	84	223468	47.714	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.420%
26) 1,2-Dichloroethane-d4	5.027	65	154587	51.180	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.360%
32) Benzene-d6	5.043	84	480587	49.927	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.860%
36) 1,2-Dichloropropane-d6	6.063	67	149939	47.775	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.560%
41) Toluene-d8	7.310	98	432442	50.178	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.360%
43) trans-1,3-Dichloroprop...	7.616	79	78544	49.100	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	98.200%
47) 2-Hexanone-d5	8.085	63	133782	78.361	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	78.360%
56) 1,1,2,2-Tetrachloroeth...	10.207	84	184424	42.943	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	85.880%
66) 1,2-Dichlorobenzene-d4	11.616	152	153844	49.626	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.260%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100822\
Data File : VV028443.D
Acq On : 08 Oct 2022 17:49
Operator : SY/MD
Sample : VV1008WBL02
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK272

Quant Time: Oct 09 05:52:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM100722WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Oct 08 02:10:35 2022
Response via : Initial Calibration

