


Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV100822\
 Data File : VV028450.D
 Acq On : 08 Oct 2022 20:37
 Operator : SY/MD
 Sample : N4889-05DL 2000X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EXZ63DL

Quant Time: Oct 09 05:53:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM100722WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Oct 08 02:10:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	318184	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	286066	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.243	152	140750	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	87405	35.601	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	71.200%
7) Chloroethane-d5	1.564	69	69474	35.688	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	71.380%
11) 1,1-Dichloroethene-d2	2.105	63	117049	28.013	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	56.020%#
21) 2-Butanone-d5	3.892	46	132187	61.505	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	61.500%
24) Chloroform-d	4.343	84	195342	43.382	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.760%
26) 1,2-Dichloroethane-d4	5.027	65	136688	47.070	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.140%
32) Benzene-d6	5.043	84	415775	45.358	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.720%
36) 1,2-Dichloropropane-d6	6.063	67	130377	43.624	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	87.240%
41) Toluene-d8	7.310	98	373942	45.564	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.120%
43) trans-1,3-Dichloroprop...	7.619	79	65306	42.870	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.740%
47) 2-Hexanone-d5	8.085	63	125443	77.159	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	77.160%
56) 1,1,2,2-Tetrachloroeth...	10.207	84	165988	40.587	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	81.180%
66) 1,2-Dichlorobenzene-d4	11.616	152	134009	45.950	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.900%
Target Compounds						
42) Toluene	7.381	91	1011811	104.334	ug/L	99
52) Ethylbenzene	9.011	91	24923	2.358	ug/L	97
53) m,p-Xylene	9.130	106	30792	7.686	ug/L	96
54) o-Xylene	9.542	106	10088	2.515	ug/L	97





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#42
Toluene
Concen: 104.334 ug/L
RT: 7.381 min Scan# 1972
Delta R.T. -0.000 min
Lab File: VV028450.D
Acq: 08 Oct 2022 20:37




Instrument :
MSVOA_V
ClientSampleId :
EXZ63DL


Tgt Ion: 91 Resp: 1011811
Ion Ratio Lower Upper
91 100
92 57.9 40.9 76.0



#52
Ethylbenzene
Concen: 2.358 ug/L
RT: 9.011 min Scan# 2479
Delta R.T. 0.006 min
Lab File: VV028450.D
Acq: 08 Oct 2022 20:37

Tgt Ion: 91 Resp: 24923
Ion Ratio Lower Upper
91 100
106 32.3 21.4 39.8









#53
m,p-Xylene
Concen: 7.686 ug/L
RT: 9.130 min Scan# 2516
Delta R.T. -0.000 min
Lab File: VV028450.D
Acq: 08 Oct 2022 20:37

Instrument :
MSVOA_V
ClientSampleId :
EXZ63DL

Tgt Ion:106 Resp: 30792
Ion Ratio Lower Upper
106 100
91 201.7 145.2 269.6



#54
o-Xylene
Concen: 2.515 ug/L
RT: 9.542 min Scan# 2644
Delta R.T. 0.006 min
Lab File: VV028450.D
Acq: 08 Oct 2022 20:37

Tgt Ion:106 Resp: 10088
Ion Ratio Lower Upper
106 100
91 212.5 152.5 283.3

