

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101322\
 Data File : VV028541.D
 Acq On : 13 Oct 2022 16:53
 Operator : SY/MD
 Sample : N5070-09ME
 Misc : 6.73g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0BQ7ME

Quant Time: Oct 14 04:51:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM100722WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Oct 14 04:49:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	247841	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	245156	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	137691	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	26087	13.641	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	27.280%#
7) Chloroethane-d5	1.564	69	25268	16.664	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	33.320%#
11) 1,1-Dichloroethene-d2	2.088	63	56073	17.228	ug/L	-0.02
Spiked Amount	50.000	Range	60 - 125	Recovery	=	34.460%#
21) 2-Butanone-d5	3.924	46	36608	21.868	ug/L	0.04
Spiked Amount	100.000	Range	40 - 130	Recovery	=	21.870%#
24) Chloroform-d	4.339	84	110183	31.415	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	62.840%#
26) 1,2-Dichloroethane-d4	5.024	65	81167	35.884	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	71.760%
32) Benzene-d6	5.040	84	171713	21.859	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	43.720%#
36) 1,2-Dichloropropane-d6	6.063	67	61978	24.198	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	48.400%#
41) Toluene-d8	7.310	98	155053	22.046	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	44.100%#
43) trans-1,3-Dichloroprop...	7.619	79	37233	28.520	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	57.040%#
47) 2-Hexanone-d5	8.101	63	28114	20.178	ug/L	0.02
Spiked Amount	100.000	Range	45 - 130	Recovery	=	20.180%#
56) 1,1,2,2-Tetrachloroeth...	10.217	84	116012	33.100	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	66.200%
66) 1,2-Dichlorobenzene-d4	11.619	152	96099	33.683	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	67.360%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101322\
Data File : VV028541.D
Acq On : 13 Oct 2022 16:53
Operator : SY/MD
Sample : N5070-09ME
Misc : 6.73g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0BQ7ME

Quant Time: Oct 14 04:51:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM100722WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Oct 14 04:49:18 2022
Response via : Initial Calibration

